

# HOMOCYSTEINE 5 LEVEL CALIBRATOR

## CALIBRATOR FOR HOMOCYSTEINE ASSAY

### Introduction

Homocysteine 5 level calibrator is intended to be used for the calibration of Homocysteine assay.

### Calibrator Storage, Stability & Handling

#### Shelf - Life

Unopened calibrators are stable when stored at 2 - 8°C until the expiration date printed on the vial label.

#### Preparation of Calibrator & Stability

Homocysteine Calibrators are liquid, ready to use. **DO NOT FREEZE.**

#### Calibrator Concentration

Concentration of each calibrator is printed on the vial label.

#### Indications of Deterioration

Presence of extreme turbidity or microbial growth may indicate deterioration.

#### Components

Homocysteine 5 level Calibrator is preparation of human serum.

#### Calibration Procedure

The calibrators should be used to analyze patient samples in accordance with the instructions outlined in Accurex Homocysteine Reagent package insert. Refer to the instrument manufacturer's recommendation for calibration frequency.

#### Precautions & Warnings

1. For *in vitro diagnostic use*.
2. Do not pipette by mouth.
3. Human material used in the preparation of this product has been tested and found nonreactive for the presence of HBsAg, HCV and antibody to HIV 1/2. However, all human-based products should still be handled in accordance with good laboratory practices using appropriate precautions to avoid transmission of any infectious diseases.
4. Do not use the calibrator after the expiration date printed on the box.
5. Sodium azide has been added to the calibrators as an antibacterial agent. Sodium azide may react with lead and copper plumbing to form potentially explosive metal azide buildup.

# HOMOCYSTEINE 5 LEVEL CALIBRATOR

## CALIBRATOR FOR HOMOCYSTEINE ASSAY

### Introduction

Homocysteine 5 level calibrator is intended to be used for the calibration of Homocysteine assay.

### Calibrator Storage, Stability & Handling

#### Shelf - Life

Unopened calibrators are stable when stored at 2 - 8°C until the expiration date printed on the vial label.

#### Preparation of Calibrator & Stability

Homocysteine Calibrators are liquid, ready to use. **DO NOT FREEZE.**

#### Calibrator Concentration

Concentration of each calibrator is printed on the vial label.

#### Indications of Deterioration

Presence of extreme turbidity or microbial growth may indicate deterioration.

#### Components

Homocysteine 5 level Calibrator is preparation of human serum.

#### Calibration Procedure

The calibrators should be used to analyze patient samples in accordance with the instructions outlined in Accurex Homocysteine Reagent package insert. Refer to the instrument manufacturer's recommendation for calibration frequency.

#### Precautions & Warnings

1. For *in vitro diagnostic use*.
2. Do not pipette by mouth.
3. Human material used in the preparation of this product has been tested and found nonreactive for the presence of HBsAg, HCV and antibody to HIV 1/2. However, all human-based products should still be handled in accordance with good laboratory practices using appropriate precautions to avoid transmission of any infectious diseases.
4. Do not use the calibrator after the expiration date printed on the box.
5. Sodium azide has been added to the calibrators as an antibacterial agent. Sodium azide may react with lead and copper plumbing to form potentially explosive metal azide buildup.



**ACCUREX BIOMEDICAL PVT. LTD.**

Head Office - Mumbai. Tel.: 91(022) 67446744

Plant : G-54/55, MIDC Tarapur, Boisar - 401 506. INDIA.

Customer Support No.: 1800 209 8456

www.accurex.net / support@accurex.net



**ACCUREX BIOMEDICAL PVT. LTD.**

Head Office - Mumbai. Tel.: 91(022) 67446744

Plant : G-54/55, MIDC Tarapur, Boisar - 401 506. INDIA.

Customer Support No.: 1800 209 8456

www.accurex.net / support@accurex.net